

Dyneon GmbH & Co.KG

Werk Gendorf
D-84504 Burgkirchen
Phone: +49-8679-7-4709
Fax: +49-8679-3992
Internet: www.dyneon.com

Stefan Ebmeyer
Product Stewardship, Fluoroplastics
Phone: +49 (8679) 7 45 07
Fax: +49 (8679) 7 91 45 07
Email: Sebmeyer@mmm.co

Burgkirchen, 21.05.2004

Dyneon™ TFM™ 1600 PTFE

USP Class VI – Confirmation of Compliance

It is hereby confirmed that representative samples manufactured with Dyneon™ TFM™ PTFE have been certified in accordance with USP protocol (USP Class VI - 70°C, Toxikon Corporation – USA, Report # 98G-2147, dated 11/06/98).

The test articles meet the requirements of USP 23, 1995, for Class VI Plastics - 70°C.

The above includes USP Class VI certification for Dyneon™ TFM™ 1600 PTFE in accordance with the above mentioned USP protocol.

We expect Dyneon™ TFM™ 1600 PTFE to successfully pass testing and certification in accordance with USP protocol based on existing evident indications and based on experience with so far completed USP tests and reports on Dyneon products provided finished articles are fabricated by good manufacturing practice and are generally suitable for the intended use.

All testing in accordance with USP protocol had to be carried out on finished products as “article tests” and is the responsibility of the manufacturer or seller.

It is the responsibility of the customer to determine whether its specific formulation and intended use comply with applicable laws and is suitable for its intended applications.

Please feel free to contact us for any further question.

Kind regards



Stefan Ebmeyer
Dyneon GmbH & Co. KG
APD Fluoroplastics